SUPERMAX
GLOVE MANUFACTURING SDN. BHD. (218698-T)

Lot 38, Putra Industrial Park, Bukit Rahman Putra,
47000 Sungai Buloh, Selangor Darul Ehsan, Malaysia.
Tel: 603-61452328 (30 lines) Fax: 603-61562191 / 67307883
http://www.supergloves.com E-MAIL: smax@po.jaring.my

USER'S MANUAL OF POWDERED LATEX EXAMINATION GLOVES

1. Identity of Manufacturer: Supermax Glove Manufacturing Sdn Bhd

2. Country of Origin: Made in Malaysia

3. Statement of Identity
   Trade Name : Supermax
   Common Name : Powdered Latex Examination Gloves

4. The "Lot Number" is preceded by a serial number.
   Example:
   Lot number – 8 24 1 7148
   Where
   8 - Supermax Glove Manufacturing Sdn Bhd
   24 - 24th week of the year
   1 - Size of the glove (0- size XS, 1- size S, 2- size M, 3- size L & 4- size XL)
   7148 - Running Number

5. Net Quantity of Contents Statement
   The label shall contain a statement of net quantity of contents in terms of numerical count. Example:-

   Quantity: 100 Gloves by Weight

6. Type of Packaging Suitable for Transport.
   - 100pieces per dispenser box
   - 10/20 dispenser boxes per carton

7. Usage of Product
   Do not require extra method of use.
   7.1 Preparation for use
   -Ensure selection of gloves with correct size.

SUPERMAX
GLOVE MANUFACTURING SDN. BHD. (218698-T)
Lot 38, Putra Industrial Park, Bukit Rahman Putra,
47000 Sungai Buloh, Selangor Darul Ehsan, Malaysia.
Tel: 603-61452328 (30 lines) Fax: 603-61562191 / 67307883
http://www.supergloves.com E-MAIL: smax@po.jaring.my

Below are the correct glove size width:

Size	XS	S	M	L	XL
Width Range	<80mm	82-88mm	92-98mm	102-108mm	>110mm

7.2 How to use

- Dry hand before donning the gloves.
- Take out the product after checking package.
- Glove need to be donned carefully; do not apply excessive force to pull the gloves when donning.
- Keep the fingernails short and change gloves frequently to avoid tearing issue.
- The gloves should be discarded with an appropriate waste disposal container after use according to local disposal regulations.

7.3 Cautions for Use

- The product is for single use.
- Users should consider the circumstances of use in deciding whether to remove any residual powder on gloves after donning. Powder can be removed by wiping gloves thoroughly with a sterile wet sponge, sterile wet tower or other effective method.
- Isolated cases of allergic reactions to latex product or powder have been reported. If you experience a reaction to this product, discontinue use immediately and seek medical help.

8. Storage Condition

The atmosphere must always be conducive to the storage of the gloves.

- Gloves before and after opened for use shall be stored in dark, cool and dry place.
- Gloves need to avoid area with excessive heat. The storage areas need to be controlled below 40ºC/ 104 ºF.
- Gloves need to be stored at the areas with good air ventilation.
- Gloves need to be shielded from direct sunlight and fluorescent lighting.
- Gloves neet to be stored far away from any energy radiating and heat generating equipment.

9. Expiration Date

The shelf life of the gloves is 3 years from the date of manufacture.